



4-Point PRO Mounting Hardware

For Aluminum & Carbon Series Back Supports

Stealth Products strives for 100% customer satisfaction. Your complete satisfaction is important. Please contact us with feedback or suggested changes that will help improve the quality and usability of these products. You may reach us at:



 Stealth Products, LLC

 104 John Kelly Drive, Burnet, TX 78611

 Phone: (512) 715-9995
 Toll Free: 1(800) 965-9229

 Fax: (512) 715-9954
 Toll Free: 1(800) 806-1225

 info@stealthproducts.com
 www.stealthproducts.com



MDSS GmbH Schiffgraben 41 30175 Hannover, Germany

CE

General

Read and understand all instructions prior to the use of the product. Failure to adhere to instructions and warnings in this document may result in property damage, injury, or death. Product misuse due to failure of the following instructions will void the warranty.

Immediately discontinue use if any function is compromised, parts are missing, loose, or shows signs of excessive wear. Consult with your supplier for repair, adjustment, or replacement.

If this document contains information you do not understand, or there are concerns about safety or operation, contact your supplier.

Important Information!

All persons responsible for fitting, adjustment, and daily use of the devices discussed in these instructions must be familiar with and understand all safety aspects of the devices mentioned. In order for our products to be used successfully, you must:

- Read and understand all instructions and warnings.
- Maintain our products according to our instructions on care and maintenance.
- Devices should be installed and adjusted by a trained technician.

Ordering Documentation

You can download additional copies of this manual on the Stealth website: *https://stlpro.site/stealth-docs* and search "*4 Point Pro*".

Installation Date:		
Dealer:		
Dealer Stamp:		

Serial Number:

Introduction

No person is authorized to alter, extend, or waive the warranties of Stealth Products, LLC.

Stealth Products warrants against failure due to defective materials or workmanship:

Covers: 2 years

Hardware: 5 years

Electronics: 3 years

Limitations

4.1

4.0

Stealth Products, LLC does not warrant damage due to, but not limited to:

- Misuse, abuse, or misapplication of products.
- Modification of product without written approval from Stealth Products, LLC.
- Any alteration or lack of serial number, where applicable, will automatically void this warranty.
- Stealth Products, LLC is liable for replacement parts only.
- Stealth Products, LLC is not liable for any incurred labor costs.

NOTICE

Any substitutions of supplied hardware void the warranty. Misuse of items and "acts of nature" (ex. Flood, tornado, etc.) are not covered by the warranty.

1.0 Customer Satisfaction	i
2.0 Important Information	ii
3.0 Introduction	iii
4.0 Warranty	iv
4.1 Limitations	iv
5.0 Table of Contents	v
6.0 Warning Labels	vi
6.1 Warning Labels	vi
6.2 Limited Liability	vi
6.3 Testing	vi
7.0 Parts and Accessories	vii
7.1 Tools Required for Installation & Adjustment	vii
7.2 Parts Required	vii
7.3 4-Point Pro Mounting Hardware Assembly	vii
7.4 4-Point PRO Torque Specification	viii
8.0 Installation Instructions	1
8.1 Hardware Installation	1
8.2 Angle Adjustment	4
8.3 Depth Adjustment	4
8.4 Width Adjustment	5
9.0 First Time Use	6
9.1 Dealer Assistance	6
9.2 User Testing	6
9.3 Conditions of Use	6
10.0 Maintenance	7
10.1 Cleaning	7
10.2 Disinfection	7

Warning Labels

6.1

Warnings are included for the safety of the user, client, operator and property. Please read and understand what the signal words **SAFETY**, **NOTICE**, **CAUTION**, **WARNING** and **DANGER** mean, how they could affect the user, those around the user, and property.

À DANGER	Identifies an imminent situation which (if not avoided) will result in severe injury , death , and property damage .
🛦 WARNING	Identifies a potential situation which (if not avoided) will result in severe injury , death , and property damage .
▲ CAUTION	Identifies a potential situation which (if not avoided) will result in minor to moderate injury , and property damage .
NOTICE	Identifies important information not related to injury, but possible property damage .
SAFETY	Indicates steps or instructions for safe practices, reminders of safe procedures, or important safety equipment that may be necessary.

Limited Liability

Stealth Products, LLC accepts no liability for personal injury or damage to property that may arise from the failure of the user or other persons to follow the recommendations, warnings, and instructions in this manual.

Testing

Initial setup and driving should be done in an open area free of obstacles until the user is fully capable of driving safely.

The product should always be tested without any person sitting in the wheelchair until every alteration of the physical installation or adjustment is complete.

vi

6.2

Par	ts and Accessories	7.0
То	ols Required for Installation & Adjustment	7.1
•	4mm Allen Wrench	
•	10mm wrench	
Ра	rts Required	7.2
•	Back Support Shell, 1 each (A)	
•	Bent Bracket Assembly, 2 each, attached to back (BR5030)(B)	
•	Inside Clamp, 4 each (BR5021)(C)	
•	Outside Clamp, 4 each (BR5022)(D)	
•	M6 x 35mm Bolt, 4 each (HBR071)(E)	
•	Nylock Nut, 4 each (HBR034)(F)	
•	M6 Fender Washer, 4 each (HBR035)(G)	
٠	M6 ID x 10mm OD x 1mm, 4 each (HBR072)(H)	
4-	Point PRO Mounting Hardware Assembly	7.3
	A	

NOTICE

В

Only use the hardware supplied with the ADI back on the ADI back. Do not substitute any other screws, brackets, or clamps. Use of alternative hardware or mounting brackets void all warranties.



Stud Plate (Inside of Shell) Use: 10mm hex socket or wrench Torque Specs: 117 in-lbs/ 13.2 N-m

Hardware Installation

- 1. Remove the existing wheelchair back according to the original manufacturer's instructions.
- 2. Visually determine the desired height of the back support.
- 3. Mark on the wheelchair back posts the desired location of mounting hardware.
- 4. Place one outer clamp (BR5022) and one inner clamp (BR5021) around the back cane on one side of the chair. Fasten with a 35mm bolt (HBR071).

NOTICE

Clamp assembly should be rotated out slightly. Do not tighten completely at this stage.



Step 3

Step 2



Step 4



- 5. Repeat step 4 on the other side of the chair.
- Hold backrest in place and turn left and right clamp assemblies until the 35mm bolts (HBR071) slide into the top slots on the bent brackets (BR5030).





▲ CAUTION


If you are able to fold the back canes, the back should be in the folded position when transporting the wheelchair, especially before loading for air travel.

- 7. Slide the remaining inner clamps (BR5021) and outer clamps (BR5022) into place with the 35mm bolt (HBR071) running through the bottom slots of the bent bracket (BR5030).
- 8. Fasten the bent brackets (BR5030) onto all four 35mm bolts (HRB071) with a washer (HBR035) and nylock nut (HBR034).



Angle Adjustment

- 1. On each side of the hardware, loosen the top and bottom nuts (HBR034) with a 10mm wrench.
- 2. Adjust back support to achieve desired user angle.
- 3. After final adjustments are complete, tighten all screws to the proper torque specification.




Depth Adjustment

- 1. On each side of the hardware, loosen the top and bottom nuts (HBR034) with a 10mm wrench.
- 2. Slide back support forward or aft to achieve desired users depth.
- 3. After final adjustments are complete, tighten all screws to the proper torque specification.



Width Adjustment

- 1. Remove fabric cover from back support.
- 2. On each side of the hardware, loosen the top and bottom nuts (as shown), using a 4mm Allen Wrench and a 10mm wrench.
- 3. Adjust hardware inward or outward to achieve required mounting hardware width.
- 4. After final adjustments are complete, tighten all screws to the proper torque specification.



M WARNING

Protect all backs from blunt forces that may cause damage and/or breakage.

▲ CAUTION

When positioning an ADI back, the interior portion of the back support should be positioned to make a contact at the PSIS to provide optimal pelvic support and trunk control.

Dealer Assistance

During first time use by the client, it is advised that the dealer or service technician assists and explains the set-up to the customer (the user and/or the attendant). If needed, the dealer can make final adjustments.

${}^{\rm th}$ caution	ADI backs may not fit every chair. Please consult your clinician or supplier providing you equipment.
▲ CAUTION	Do not attach the ADI backs with the user in the chair.

User Testing

Its is important that the customer is fully aware of the installation, how to use it, and what can be adjusted in order to gain as much mobility as possible. As a dealer, proceed as follows:

- Explain and show the customer how you have executed the installation, and explain the function of the hardware.
- If needed, adjust the hardware to the proper position.
- Explain to the customer possible problems and how to address them.
- •

CAUTION Do not use ADI backs during vehicle transportation.

Conditions of Use

The ADI hardware is intended for use as installed by the dealer, in accordance to the installation instructions in this manual.

- The foreseen conditions of use are communicated by the dealer or service technician to the user and/or attendant during the first time use.
- If the usage conditions change significantly, please contact your dealer or a qualified service technician to avoid excessive wear and tear or unintended damage.

9.1

9.3

Cleaning

10.1

- Use a soft cloth to clean the back shell.
- Hand wash the foam pad cover with mild, soapy water or detergent and let dry.
- Foam pad should NOT get wet.

WARNING Do not hand wash in water temperatures above 160°F (70° C).

▲ CAUTION

Institution MUST clean and disinfect product between multiple patient uses.

🛦 WARNING	Do not wash or machine dry.
-----------	-----------------------------

Disinfection

10.2

- Wipe gently with a cloth dampened with a household disinfectant.
- Allow foam to air dry thoroughly.

A WARNING Do not immerse the foam in water.

SAFETY

Always check all mounting hardware, making sure each fastener is properly tightened before using the back support.

A WARNING 250 lbs. (113.40 kg) user weight limit.

NOTICE

Altering an ADI back in any way and/or improper installation will void the warranty.

ΝI	0		0	0
IN	υ	U	E	S

INI	otac
	ULES

INI	\sim	11	(a)	[9]
	U	u	-	2



